

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028729.D
 Acq On : 16 May 2022 14:21
 Operator : JC/MD
 Sample : N2862-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2R

Quant Time: May 17 03:36:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	230419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	424238	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	407314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166379	52.676	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 105.360%		
35) Dibromofluoromethane	5.391	113	141606	50.946	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 101.900%		
50) Toluene-d8	8.653	98	512451	48.901	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 97.800%		
62) 4-Bromofluorobenzene	11.079	95	189215	48.342	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 96.680%		
Target Compounds						Qvalue
16) Acetone	2.380	43	3391	2.502	ug/l	90
30) Chloroform	5.105	83	1739	0.338	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028729.D
Acq On : 16 May 2022 14:21
Operator : JC/MD
Sample : N2862-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2R

Quant Time: May 17 03:36:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

