

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028731.D
 Acq On : 16 May 2022 15:08
 Operator : JC/MD
 Sample : N2862-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4R

Quant Time: May 17 03:36:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	245953	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	447504	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	429843	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	189375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	174562	51.777	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.560%
35) Dibromofluoromethane	5.391	113	148974	50.810	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.620%
50) Toluene-d8	8.653	98	538794	48.742	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.480%
62) 4-Bromofluorobenzene	11.085	95	197253	47.775	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.560%
Target Compounds						
						Qvalue
7) Trichlorofluoromethane	1.892	101	1777	0.341	ug/l	91
16) Acetone	2.380	43	2966	2.050	ug/l	# 87
30) Chloroform	5.105	83	2494	0.454	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028731.D
Acq On : 16 May 2022 15:08
Operator : JC/MD
Sample : N2862-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4R

Quant Time: May 17 03:36:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

