

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028734.D
 Acq On : 16 May 2022 16:18
 Operator : JC/MD
 Sample : N2862-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: May 17 03:38:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	270419	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	524460	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	436224	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	195053	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	195399	52.713	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.420%	
35) Dibromofluoromethane	5.385	113	153475	44.664	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 89.320%	
50) Toluene-d8	8.653	98	555773	42.900	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 85.800%	
62) 4-Bromofluorobenzene	11.079	95	204171	42.195	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 84.380%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028734.D
Acq On : 16 May 2022 16:18
Operator : JC/MD
Sample : N2862-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: May 17 03:38:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

