

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028738.D
 Acq On : 16 May 2022 17:52
 Operator : JC/MD
 Sample : PB144860TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144860TB

Quant Time: May 17 03:39:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	253561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	460852	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	434760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	199234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	178084	51.236	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.480%	
35) Dibromofluoromethane	5.391	113	151718	50.247	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.500%	
50) Toluene-d8	8.653	98	549702	48.288	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.580%	
62) 4-Bromofluorobenzene	11.079	95	206165	48.487	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.980%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	3761	1.202	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028738.D
Acq On : 16 May 2022 17:52
Operator : JC/MD
Sample : PB144860TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144860TB

Quant Time: May 17 03:39:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

