

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028742.D
 Acq On : 16 May 2022 19:26
 Operator : JC/MD
 Sample : N2867-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW205S1-250-252

Quant Time: May 17 03:41:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	224944	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	420862	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	407588	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184687	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161542	52.390	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.780%	
35) Dibromofluoromethane	5.385	113	137978	50.039	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.080%	
50) Toluene-d8	8.653	98	569695	54.799	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 109.600%	
62) 4-Bromofluorobenzene	11.079	95	191282	49.262	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.520%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	8484	6.411	ug/l	96
17) Carbon Disulfide	2.514	76	2203	0.365	ug/l	96
25) 2-Butanone	4.574	43	2427	1.168	ug/l #	88
44) Trichloroethene	7.129	130	3654	1.132	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028742.D
Acq On : 16 May 2022 19:26
Operator : JC/MD
Sample : N2867-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW205S1-250-252

Quant Time: May 17 03:41:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

