

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028745.D
 Acq On : 16 May 2022 20:36
 Operator : JC/MD
 Sample : N2806-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-8/9-1

Quant Time: May 17 03:42:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	261080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	475227	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	460856	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	216926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	183121	51.168	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.340%	
35) Dibromofluoromethane	5.391	113	153840	49.409	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.820%	
50) Toluene-d8	8.653	98	575978	49.066	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.140%	
62) 4-Bromofluorobenzene	11.079	95	223145	50.893	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.780%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	64764	42.166	ug/l	92
20) Methylene Chloride	2.788	84	40925	12.700	ug/l	93
25) 2-Butanone	4.568	43	16627	6.897	ug/l	98
31) Cyclohexane	5.471	56	7304	1.470	ug/l #	89
39) Methylcyclohexane	7.385	83	71880	13.403	ug/l	97
43) Isopropyl Acetate	6.348	43	24913	3.260	ug/l #	92
65) Chlorobenzene	10.079	112	54238	5.548	ug/l	98
73) Isopropylbenzene	10.963	105	62928	3.825	ug/l	98
78) n-propylbenzene	11.305	91	47611	2.502	ug/l	98
85) sec-Butylbenzene	11.890	105	29979	1.831	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028745.D
Acq On : 16 May 2022 20:36
Operator : JC/MD
Sample : N2806-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-8/9-1

Quant Time: May 17 03:42:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

