

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028719.D
 Acq On : 16 May 2022 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 17 03:31:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.655	0.622	5.0	90	0.00
3 P	Chloromethane	0.617	0.577	6.5	95	0.00
4 C	Vinyl Chloride	0.664	0.632	4.8#	92	0.00
5 T	Bromomethane	0.523	0.444	15.1	89	0.00
6 T	Chloroethane	0.467	0.425	9.0	95	0.00
7 T	Trichlorofluoromethane	1.059	1.025	3.2	92	0.00
8 T	Diethyl Ether	0.386	0.417	-8.0	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.551	1.1	97	0.00
10 T	Methyl Iodide	0.667	0.552	17.2	77	0.00
11 T	Tert butyl alcohol	0.156	0.129	17.3	89	0.00
12 CM	1,1-Dichloroethene	0.517	0.502	2.9#	94	0.00
13 T	Acrolein	0.112	0.118	-5.4	114	0.00
14 T	Allyl chloride	0.830	0.838	-1.0	97	0.00
15 T	Acrylonitrile	0.330	0.321	2.7	94	0.00
16 T	Acetone	0.294	0.309	-5.1	107	0.00
17 T	Carbon Disulfide	1.341	1.326	1.1	93	0.00
18 T	Methyl Acetate	0.737	0.716	2.8	97	0.00
19 T	Methyl tert-butyl Ether	1.836	1.838	-0.1	96	0.00
20 T	Methylene Chloride	0.617	0.582	5.7	95	0.00
21 T	trans-1,2-Dichloroethene	0.583	0.562	3.6	91	0.00
22 T	Diisopropyl ether	1.774	1.791	-1.0	96	0.00
23 T	Vinyl Acetate	1.529	1.578	-3.2	96	0.00
24 P	1,1-Dichloroethane	1.031	1.023	0.8	95	0.00
25 T	2-Butanone	0.462	0.460	0.4	97	0.00
26 T	2,2-Dichloropropane	0.808	0.855	-5.8	99	0.00
27 T	cis-1,2-Dichloroethene	0.681	0.659	3.2	94	0.00
28 T	Bromochloromethane	0.434	0.456	-5.1	117	0.00
29 T	Tetrahydrofuran	0.300	0.289	3.7	93	0.00
30 C	Chloroform	1.116	1.092	2.2#	93	0.00
31 T	Cyclohexane	0.951	0.934	1.8	94	0.00
32 T	1,1,1-Trichloroethane	0.974	0.967	0.7	93	0.00
33 S	1,2-Dichloroethane-d4	0.685	0.661	3.5	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.328	0.328	0.0	95	0.00
36 T	1,1-Dichloropropene	0.492	0.495	-0.6	94	0.00
37 T	Ethyl Acetate	0.514	0.525	-2.1	93	0.00
38 T	Carbon Tetrachloride	0.495	0.520	-5.1	95	0.00
39 T	Methylcyclohexane	0.564	0.617	-9.4	100	0.00
40 TM	Benzene	1.414	1.409	0.4	93	0.00
41 T	Methacrylonitrile	0.276	0.283	-2.5	98	0.00
42 TM	1,2-Dichloroethane	0.522	0.525	-0.6	95	0.00
43 T	Isopropyl Acetate	0.804	0.824	-2.5	95	0.00
44 TM	Trichloroethene	0.384	0.381	0.8	91	0.00
45 C	1,2-Dichloropropane	0.353	0.354	-0.3#	95	0.00
46 T	Dibromomethane	0.263	0.266	-1.1	95	0.00
47 T	Bromodichloromethane	0.489	0.518	-5.9	95	0.00
48 T	Methyl methacrylate	0.388	0.408	-5.2	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028719.D
 Acq On : 16 May 2022 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 17 03:31:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	88	0.00
50 S	Toluene-d8	1.235	1.225	0.8	92	0.00
51 T	4-Methyl-2-Pentanone	0.530	0.540	-1.9	94	0.00
52 CM	Toluene	0.910	0.924	-1.5#	93	0.00
53 T	t-1,3-Dichloropropene	0.496	0.550	-10.9	95	0.00
54 T	cis-1,3-Dichloropropene	0.550	0.595	-8.2	95	0.00
55 T	1,1,2-Trichloroethane	0.371	0.378	-1.9	94	0.00
56 T	Ethyl methacrylate	0.562	0.587	-4.4	94	0.00
57 T	1,3-Dichloropropane	0.626	0.628	-0.3	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.275	0.325	-18.2	112	0.00
59 T	2-Hexanone	0.411	0.422	-2.7	93	0.00
60 T	Dibromochloromethane	0.362	0.398	-9.9	96	0.00
61 T	1,2-Dibromoethane	0.390	0.399	-2.3	94	0.00
62 S	4-Bromofluorobenzene	0.461	0.520	-12.8	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.378	0.381	-0.8	92	0.00
65 PM	Chlorobenzene	1.061	1.074	-1.2	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.376	0.391	-4.0	94	0.00
67 C	Ethyl Benzene	1.901	1.947	-2.4#	92	0.00
68 T	m/p-Xylenes	0.740	0.761	-2.8	92	0.00
69 T	o-Xylene	0.729	0.744	-2.1	93	0.00
70 T	Styrene	1.183	1.251	-5.7	92	0.00
71 P	Bromoform	0.274	0.311	-13.5	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.792	3.428	9.6	93	0.00
74 T	N-amyl acetate	1.491	1.385	7.1	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.301	1.248	4.1	103	0.00
76 T	1,2,3-Trichloropropane	1.171	1.148	2.0	102	0.00
77 T	Bromobenzene	0.897	0.883	1.6	101	0.00
78 T	n-propylbenzene	4.386	4.429	-1.0	102	0.00
79 T	2-Chlorotoluene	2.688	2.619	2.6	102	0.00
80 T	1,3,5-Trimethylbenzene	3.159	3.209	-1.6	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.357	0.346	3.1	97	0.00
82 T	4-Chlorotoluene	3.108	3.075	1.1	102	0.00
83 T	tert-Butylbenzene	3.059	3.215	-5.1	107	0.00
84 T	1,2,4-Trimethylbenzene	3.168	3.272	-3.3	104	0.00
85 T	sec-Butylbenzene	3.773	4.038	-7.0	108	0.00
86 T	p-Isopropyltoluene	3.166	3.464	-9.4	108	0.00
87 T	1,3-Dichlorobenzene	1.760	1.857	-5.5	109	0.00
88 T	1,4-Dichlorobenzene	1.825	1.764	3.3	102	0.00
89 T	n-Butylbenzene	2.743	3.124	-13.9	110	0.00
90 T	Hexachloroethane	0.512	0.589	-15.0	112	0.00
91 T	1,2-Dichlorobenzene	1.729	1.759	-1.7	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.278	0.260	6.5	96	0.00
93 T	1,2,4-Trichlorobenzene	0.998	0.948	5.0	96	0.00
94 T	Hexachlorobutadiene	0.406	0.413	-1.7	103	0.00
95 T	Naphthalene	3.553	3.346	5.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028719.D
Acq On : 16 May 2022 10:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 17 03:31:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.003	0.941	6.2	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6