

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
 Data File : VX035686.D
 Acq On : 16 May 2023 13:20
 Operator : JC/MD
 Sample : 02751-06DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-02-71-05082023DL

Quant Time: May 17 05:29:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/17/2023
 Supervised By : Mahesh Dadoda 05/17/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	202819	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	368397	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	334865	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	167415	47.316	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.640%		
35) Dibromofluoromethane	5.385	113	118984	47.891	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.780%		
50) Toluene-d8	8.647	98	443613	48.333	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.660%		
62) 4-Bromofluorobenzene	11.079	95	181348	49.072	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.140%		
Target Compounds					Qvalue	
44) Trichloroethene	7.123	130	80975	29.109	ug/l	91
64) Tetrachloroethene	9.275	164	3921	1.739	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
Data File : VX035686.D
Acq On : 16 May 2023 13:20
Operator : JC/MD
Sample : 02751-06DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-02-71-05082023DL

Quant Time: May 17 05:29:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/17/2023
Supervised By : Mahesh Dadoda 05/17/2023

