

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
 Data File : VX035687.D
 Acq On : 16 May 2023 13:43
 Operator : JC/MD
 Sample : 02643-10DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-07-99-05032023DL

Quant Time: May 17 05:29:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	167354	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	303668	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	271575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140918	48.267	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.540%
35) Dibromofluoromethane	5.385	113	97830	47.770	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.540%
50) Toluene-d8	8.647	98	365819	48.353	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.700%
62) 4-Bromofluorobenzene	11.079	95	145755	47.848	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.700%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.123	130	96487	42.079	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
Data File : VX035687.D
Acq On : 16 May 2023 13:43
Operator : JC/MD
Sample : 02643-10DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-07-99-05032023DL

Quant Time: May 17 05:29:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

