

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
 Data File : VX035691.D
 Acq On : 16 May 2023 15:16
 Operator : JC/MD
 Sample : 02751-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-02-81-05082023

Quant Time: May 17 05:31:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	184353	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	341810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	312656	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156290	48.596	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.200%
35) Dibromofluoromethane	5.385	113	109509	47.506	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.020%
50) Toluene-d8	8.647	98	409409	48.076	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.160%
62) 4-Bromofluorobenzene	11.079	95	169367	49.395	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.780%
Target Compounds						
44) Trichloroethene	7.129	130	10858	4.207	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
Data File : VX035691.D
Acq On : 16 May 2023 15:16
Operator : JC/MD
Sample : 02751-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-02-81-05082023

Quant Time: May 17 05:31:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

