

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
 Data File : VX035692.D
 Acq On : 16 May 2023 15:39
 Operator : JC/MD
 Sample : 02751-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-01-99-05062023

Quant Time: May 17 05:31:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	178406	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	331679	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	306473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144128	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	153797	49.415	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.820%
35) Dibromofluoromethane	5.391	113	107437	48.031	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.060%
50) Toluene-d8	8.647	98	400729	48.494	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.980%
62) 4-Bromofluorobenzene	11.079	95	168851	50.749	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.500%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.319	96	3540	1.782	ug/l	98
16) Acetone	2.392	43	5938	4.439	ug/l	92
44) Trichloroethene	7.135	130	1473	0.588	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
Data File : VX035692.D
Acq On : 16 May 2023 15:39
Operator : JC/MD
Sample : 02751-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-01-99-05062023

Quant Time: May 17 05:31:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

