

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
 Data File : VX035694.D
 Acq On : 16 May 2023 16:26
 Operator : JC/MD
 Sample : 02615-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002-1

Quant Time: May 17 05:32:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	175018	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	321744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	292152	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	147740	48.388	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.780%
35) Dibromofluoromethane	5.385	113	104792	48.295	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.647	98	386193	48.178	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.360%
62) 4-Bromofluorobenzene	11.079	95	158876	49.225	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.460%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	5290	4.031	ug/l	90
30) Chloroform	5.099	83	53152	11.863	ug/l	97
44) Trichloroethene	7.129	130	490	0.202	ug/l	58
47) Bromodichloromethane	7.824	83	21768	6.484	ug/l	93
52) Toluene	8.720	92	1558	0.258	ug/l	87
60) Dibromochloromethane	9.518	129	6961	3.074	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
Data File : VX035694.D
Acq On : 16 May 2023 16:26
Operator : JC/MD
Sample : 02615-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
002-1

Quant Time: May 17 05:32:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

