

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
 Data File : VX035696.D
 Acq On : 16 May 2023 17:12
 Operator : JC/MD
 Sample : 02615-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-2

Quant Time: May 17 05:33:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	175878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	325920	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	294961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	147800	48.171	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.340%
35) Dibromofluoromethane	5.385	113	104961	47.753	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.500%
50) Toluene-d8	8.647	98	391998	48.276	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.560%
62) 4-Bromofluorobenzene	11.079	95	159923	48.915	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.820%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	14523	11.013	ug/l	89
30) Chloroform	5.092	83	16046	3.564	ug/l	91
47) Bromodichloromethane	7.824	83	2857	0.840	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
Data File : VX035696.D
Acq On : 16 May 2023 17:12
Operator : JC/MD
Sample : 02615-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-2

Quant Time: May 17 05:33:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

