

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
 Data File : VX035700.D
 Acq On : 16 May 2023 18:46
 Operator : JC/MD
 Sample : 02615-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002-3

Quant Time: May 17 05:34:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	172832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	320820	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	289926	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129219	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148487	49.247	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.500%
35) Dibromofluoromethane	5.385	113	105574	48.796	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.600%
50) Toluene-d8	8.647	98	390088	48.804	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.600%
62) 4-Bromofluorobenzene	11.079	95	157706	49.003	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.000%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	5758	4.443	ug/l	92
30) Chloroform	5.093	83	59532	13.455	ug/l	94
47) Bromodichloromethane	7.824	83	29587	8.838	ug/l	100
52) Toluene	8.720	92	1927	0.320	ug/l	100
60) Dibromochloromethane	9.525	129	9137	4.047	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
Data File : VX035700.D
Acq On : 16 May 2023 18:46
Operator : JC/MD
Sample : 02615-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
002-3

Quant Time: May 17 05:34:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

