



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051623\
Data File : VX035678.D
Acq On : 16 May 2023 09:10
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Title : SW846 8260
Last Update : Thu May 11 13:57:54 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.0	14824	PASS
75	95	30	60	55.4	37248	PASS
95	95	100	100	100.0	67235	PASS
96	95	5	9	6.9	4623	PASS
173	174	0.00	2	1.0	429	PASS
174	95	50	100	65.4	43976	PASS
175	174	5	9	7.0	3094	PASS
176	174	95	101	95.9	42179	PASS
177	176	5	9	6.5	2753	PASS