

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051624\
 Data File : VX041553.D
 Acq On : 16 May 2024 10:54
 Operator : JC/MD
 Sample : VX0516WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0516WBL01

Quant Time: May 17 02:36:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	27764	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	155536	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	137880	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	66495	30.413	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.367%
60) 4-Bromofluorobenzene	11.079	95	60101	27.923	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.067%
63) Toluene-d8	8.647	98	177891	30.098	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.333%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051624\
Data File : VX041553.D
Acq On : 16 May 2024 10:54
Operator : JC/MD
Sample : VX0516WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0516WBL01

Quant Time: May 17 02:36:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 16 05:13:13 2024
Response via : Initial Calibration

