

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051624\
 Data File : VX041556.D
 Acq On : 16 May 2024 12:25
 Operator : JC/MD
 Sample : P2503-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240507034-02

Quant Time: May 17 02:38:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.891	128	36193	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	199904	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	182222	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	84428	29.622	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.733%
60) 4-Bromofluorobenzene	11.079	95	79557	27.968	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.233%
63) Toluene-d8	8.647	98	232386	29.751	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.167%
Target Compounds						
					Qvalue	
12) Methyl Acetate	2.703	43	2919	1.047	ug/l	98
15) Acetone	2.386	58	879	2.676	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051624\
Data File : VX041556.D
Acq On : 16 May 2024 12:25
Operator : JC/MD
Sample : P2503-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240507034-02

Quant Time: May 17 02:38:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 16 05:13:13 2024
Response via : Initial Calibration

