

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051624\
 Data File : VX041550.D
 Acq On : 16 May 2024 09:09
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: May 17 02:30:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	107	0.00
2 M	Dichlorodifluoromethane	1.569	1.583	-0.9	117	0.00
3 M	Chloromethane	1.718	1.617	5.9	102	0.00
4 M	Vinyl Chloride	1.809	1.569	13.3	95	0.00
5 M	Bromomethane	1.151	1.046	9.1	103	0.00
6 M	Chloroethane	0.915	0.884	3.4	108	0.00
7 M	Trichlorofluoromethane	2.746	2.957	-7.7	117	0.00
8 T	Diethyl Ether	1.025	0.989	3.5	109	0.00
9	1,1,2-Trichlorotrifluoroeth	1.592	1.716	-7.8	109	0.00
10 M	1,1-Dichloroethene	1.535	1.565	-2.0	103	0.00
11	Methyl Iodide	1.828	1.994	-9.1	119	0.00
12	Methyl Acetate	2.312	2.285	1.2	106	0.00
13 M	Acrolein	0.281	0.275	2.1	103	0.00
14 M	Acrylonitrile	1.016	0.962	5.3	106	0.00
15 M	Acetone	0.272	0.329	-21.0	124	0.00
16 M	Carbon Disulfide	3.465	3.309	4.5	109	0.00
17	Allyl chloride	2.724	2.686	1.4	106	0.00
18 M	Methylene Chloride	1.826	1.818	0.4	105	0.00
19 M	trans-1,2-Dichloroethene	1.760	1.643	6.6	103	0.00
20 T	Diisopropyl ether	5.723	5.504	3.8	105	0.00
21 M	1,1-Dichloroethane	3.212	3.067	4.5	106	0.00
22 M	cis-1,2-Dichloroethene	2.139	2.013	5.9	106	0.00
23 M	tert-Butyl Alcohol	0.431	0.400	7.2	104	0.00
24 M	Methyl tert-Butyl Ether	5.730	5.409	5.6	105	0.00
25 M	Chloroform	3.422	3.277	4.2	106	0.00
26	Cyclohexane	2.678	2.568	4.1	106	0.00
27 s	1,2-Dichloroethane-d4	2.362	2.459	-4.1	111	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	110	0.00
29	1,1-Dichloropropene	0.397	0.376	5.3	107	0.00
30 M	2-Butanone	0.263	0.262	0.4	112	0.00
31	2,2-Dichloropropane	0.466	0.439	5.8	106	0.00
32 M	1,1,1-Trichloroethane	0.532	0.491	7.7	105	0.00
33 M	Carbon Tetrachloride	0.456	0.425	6.8	107	-0.01
34 M	Benzene	1.253	1.193	4.8	106	0.00
35	Methacrylonitrile	0.266	0.251	5.6	106	0.00
36 M	1,2-Dichloroethane	0.468	0.451	3.6	108	0.00
37 M	Trichloroethene	0.349	0.324	7.2	105	0.00
38	Methylcyclohexane	0.504	0.468	7.1	104	0.00
39 M	1,2-Dichloropropane	0.296	0.283	4.4	105	0.00
40	Dibromomethane	0.233	0.218	6.4	108	0.00
41 M	Bromodichloromethane	0.438	0.401	8.4	107	0.00
42 M	Vinyl Acetate	0.914	0.862	5.7	106	0.00
43	Ethyl Acetate	0.508	0.476	6.3	106	0.00
44	Isopropyl Acetate	0.768	0.701	8.7	104	0.00
45 T	1,4-Dioxane	0.009	0.008	11.1	105	0.00
46	Methyl methacrylate	0.382	0.353	7.6	106	0.00
47	n-amyl Acetate	0.697	0.640	8.2	106	0.00
48 M	t-1,3-Dichloropropene	0.440	0.388	11.8	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051624\
 Data File : VX041550.D
 Acq On : 16 May 2024 09:09
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: May 17 02:30:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.476	0.433	9.0	107	0.00
50 M	1,1,2-Trichloroethane	0.324	0.301	7.1	105	0.00
51	Ethyl methacrylate	0.522	0.469	10.2	105	0.00
52	1,3-Dichloropropane	0.536	0.502	6.3	104	0.00
53 M	Dibromochloromethane	0.369	0.341	7.6	110	0.00
54 M	1,2-Dibromoethane	0.344	0.330	4.1	107	0.00
55 M	2-Chloroethyl vinyl ether	0.255	0.249	2.4	108	0.00
56 M	Bromoform	0.296	0.265	10.5	112	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	111	0.00
58 M	4-Methyl-2-Pentanone	0.548	0.533	2.7	107	0.00
59 M	2-Hexanone	0.437	0.421	3.7	109	0.00
60 S	4-Bromofluorobenzene	0.468	0.482	-3.0	112	0.00
61 M	Tetrachloroethene	0.367	0.346	5.7	107	0.00
62 M	Toluene	1.493	1.411	5.5	106	0.00
63 S	Toluene-d8	1.286	1.298	-0.9	110	0.00
64 M	Chlorobenzene	1.014	0.961	5.2	108	0.00
65	1,1,1,2-Tetrachloroethane	0.359	0.334	7.0	107	0.00
66 M	Ethyl Benzene	1.658	1.549	6.6	105	0.00
67 M	m/p-Xylenes	0.648	0.627	3.2	106	0.00
68 M	o-Xylene	0.652	0.617	5.4	108	0.00
69 M	Styrene	1.094	1.021	6.7	107	0.00
70	Isopropylbenzene	1.649	1.593	3.4	106	0.00
71 M	1,1,2,2-Tetrachloroethane	0.566	0.540	4.6	108	0.00
72	1,2,3-Trichloropropane	0.511	0.501	2.0	109	0.00
73	Bromobenzene	0.484	0.456	5.8	108	0.00
74	n-propylbenzene	1.859	1.783	4.1	108	0.00
75	2-Chlorotoluene	1.167	1.112	4.7	107	0.00
76	1,3,5-Trimethylbenzene	1.407	1.333	5.3	107	0.00
77	t-1,4-Dichloro-2-butene	0.161	0.143	11.2	110	0.00
78	4-Chlorotoluene	1.330	1.262	5.1	107	0.00
79	tert-butylbenzene	1.498	1.411	5.8	107	0.00
80	1,2,4-Trimethylbenzene	1.411	1.335	5.4	106	0.00
81	sec-Butylbenzene	1.750	1.667	4.7	108	0.00
82	p-Isopropyltoluene	1.532	1.461	4.6	109	0.00
83 M	1,3-Dichlorobenzene	0.869	0.820	5.6	109	0.00
84 M	1,4-Dichlorobenzene	0.878	0.826	5.9	109	0.00
85	n-Butylbenzene	1.233	1.140	7.5	110	0.00
86 T	Hexachloroethane	0.258	0.231	10.5	109	0.00
87 M	1,2-Dichlorobenzene	0.887	0.829	6.5	107	0.00
88	1,2-Dibromo-3-Chloropropane	0.124	0.112	9.7	107	0.00
89	1,2,4-Trichlorobenzene	0.596	0.556	6.7	110	0.00
90	Hexachlorobutadiene	0.291	0.275	5.5	109	0.00
91 M	Naphthalene	1.781	1.652	7.2	108	0.00
92	1,2,3-Trichlorobenzene	0.609	0.580	4.8	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0