

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051625\
 Data File : VX046235.D
 Acq On : 16 May 2025 12:11
 Operator : JC/MD
 Sample : Q2034-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 L3-WC-6

Quant Time: May 16 23:03:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124293	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	118028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50577	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63218	54.293	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.580%	
35) Dibromofluoromethane	5.385	113	46198	51.616	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.240%	
50) Toluene-d8	8.647	98	159543	51.501	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.000%	
62) 4-Bromofluorobenzene	11.079	95	62004	52.179	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.360%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	18385	39.380	ug/l	97
18) Methyl Acetate	2.709	43	2122	1.959	ug/l #	87
20) Methylene Chloride	2.788	84	18393	20.559	ug/l	96
25) 2-Butanone	4.580	43	5822	8.590	ug/l #	86
43) Isopropyl Acetate	6.342	43	13394	5.909	ug/l	97
52) Toluene	8.720	92	2459	1.138	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051625\
Data File : VX046235.D
Acq On : 16 May 2025 12:11
Operator : JC/MD
Sample : Q2034-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
L3-WC-6

Quant Time: May 16 23:03:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

