

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051625\
 Data File : VX046237.D
 Acq On : 16 May 2025 12:58
 Operator : JC/MD
 Sample : Q2048-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 L2-WC-2

Quant Time: May 16 23:04:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63018	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127745	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	124101	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55956	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64737	55.102	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.200%
35) Dibromofluoromethane	5.385	113	46411	50.452	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.900%
50) Toluene-d8	8.647	98	163100	51.227	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.460%
62) 4-Bromofluorobenzene	11.079	95	64445	52.768	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.540%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	17687	37.547	ug/l	99
18) Methyl Acetate	2.709	43	2671	2.443	ug/l	94
20) Methylene Chloride	2.788	84	11857	13.135	ug/l	91
25) 2-Butanone	4.587	43	5553	8.120	ug/l	97
43) Isopropyl Acetate	6.342	43	13904	5.968	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051625\
Data File : VX046237.D
Acq On : 16 May 2025 12:58
Operator : JC/MD
Sample : Q2048-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
L2-WC-2

Quant Time: May 16 23:04:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

