

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051625\
 Data File : VX046239.D
 Acq On : 16 May 2025 13:45
 Operator : JC/MD
 Sample : Q2048-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 L2-WC-4

Quant Time: May 16 23:05:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125976	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63360	53.706	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.420%	
35) Dibromofluoromethane	5.385	113	46208	50.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.880%	
50) Toluene-d8	8.647	98	163082	51.940	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.880%	
62) 4-Bromofluorobenzene	11.079	95	64277	53.369	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.740%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	18899	39.953	ug/l	100
18) Methyl Acetate	2.709	43	3002	2.735	ug/l	96
20) Methylene Chloride	2.788	84	13448	14.835	ug/l	96
25) 2-Butanone	4.580	43	5928	8.632	ug/l #	82
43) Isopropyl Acetate	6.348	43	13857	6.032	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051625\
Data File : VX046239.D
Acq On : 16 May 2025 13:45
Operator : JC/MD
Sample : Q2048-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
L2-WC-4

Quant Time: May 16 23:05:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

