

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009655.D
 Acq On : 17 May 2019 11:46
 Operator : JC/SP
 Sample : K2861-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20190514

Quant Time: May 20 11:22:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X050119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 02 06:46:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	26158	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	148738	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	126979	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.06	65	66120	30.47	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.57%
60) 4-Bromofluorobenzene	11.13	95	59797	27.71	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.37%
63) Toluene-d8	8.71	98	181583	30.16	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.53%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009655.D
Acq On : 17 May 2019 11:46
Operator : JC/SP
Sample : K2861-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20190514

Quant Time: May 20 11:22:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X050119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 02 06:46:57 2019
Response via : Initial Calibration

