

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009679.D
 Acq On : 17 May 2019 22:03
 Operator : JC/SP
 Sample : K2891-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-GW-578-580

Quant Time: May 18 06:32:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	157286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	251116	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	214769	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	81876	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	107308	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
35) Dibromofluoromethane	5.50	113	77364	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
50) Toluene-d8	8.71	98	315540	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
62) 4-Bromofluorobenzene	11.13	95	98108	42.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.72%	
Target Compounds						
16) Acetone	2.45	43	5579	4.839	ug/l	97
52) Toluene	8.79	92	4127	0.940	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009679.D
Acq On : 17 May 2019 22:03
Operator : JC/SP
Sample : K2891-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB202-GW-578-580

Quant Time: May 18 06:32:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

