

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009680.D
 Acq On : 17 May 2019 22:27
 Operator : JC/SP
 Sample : K2891-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-GW-603-605

Quant Time: May 18 06:33:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	163471	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	260019	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	227055	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	92647	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	110123	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
35) Dibromofluoromethane	5.50	113	80652	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
50) Toluene-d8	8.71	98	326996	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
62) 4-Bromofluorobenzene	11.13	95	104813	43.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.42%	
Target Compounds						
16) Acetone	2.45	43	8591	7.170	ug/l	97
52) Toluene	8.79	92	15632	3.439	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009680.D
Acq On : 17 May 2019 22:27
Operator : JC/SP
Sample : K2891-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB202-GW-603-605

Quant Time: May 18 06:33:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

