

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009682.D
 Acq On : 17 May 2019 23:14
 Operator : JC/SP
 Sample : K2887-22
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW726-190515

Quant Time: May 18 06:36:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	159339	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	255201	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	222017	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	88637	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	107581	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
35) Dibromofluoromethane	5.50	113	78847	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
50) Toluene-d8	8.71	98	319956	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
62) 4-Bromofluorobenzene	11.13	95	101746	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	3630	1.876	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009682.D
Acq On : 17 May 2019 23:14
Operator : JC/SP
Sample : K2887-22
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW726-190515

Quant Time: May 18 06:36:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

