

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009684.D
 Acq On : 18 May 2019 00:01
 Operator : JC/SP
 Sample : K2892-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20190515

Quant Time: May 18 06:40:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	158805	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	250858	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	218617	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	84126	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107117	49.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.82%	
35) Dibromofluoromethane	5.50	113	78555	49.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.62%	
50) Toluene-d8	8.71	98	317826	49.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.84%	
62) 4-Bromofluorobenzene	11.13	95	99607	43.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009684.D
Acq On : 18 May 2019 00:01
Operator : JC/SP
Sample : K2892-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20190515

Quant Time: May 18 06:40:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

