

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009690.D
 Acq On : 18 May 2019 03:32
 Operator : JC/SP
 Sample : VX0517WBL03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0517WBL03

Quant Time: May 18 06:49:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	165345	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	266302	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	230143	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	95901	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120822	53.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.06%	
35) Dibromofluoromethane	5.50	113	87956	52.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.08%	
50) Toluene-d8	8.71	98	357662	52.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.84%	
62) 4-Bromofluorobenzene	11.13	95	115281	46.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009690.D
Acq On : 18 May 2019 03:32
Operator : JC/SP
Sample : VX0517WBL03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0517WBL03

Quant Time: May 18 06:49:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

