

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009693.D
 Acq On : 18 May 2019 04:42
 Operator : JC/SP
 Sample : K2901-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-051519

Quant Time: May 18 06:54:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	169004	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	268694	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	229957	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	89882	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	112110	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
35) Dibromofluoromethane	5.50	113	81570	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
50) Toluene-d8	8.71	98	336516	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
62) 4-Bromofluorobenzene	11.13	95	105634	42.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.26%	
Target Compounds						
16) Acetone	2.45	43	9919	8.007	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009693.D
Acq On : 18 May 2019 04:42
Operator : JC/SP
Sample : K2901-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-051519

Quant Time: May 18 06:54:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

