

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009698.D
 Acq On : 18 May 2019 06:39
 Operator : JC/SP
 Sample : K2901-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10B_R2

Quant Time: May 20 04:26:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	168449	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	267469	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	233136	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	94722	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	113291	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
35) Dibromofluoromethane	5.50	113	82498	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
50) Toluene-d8	8.71	98	334895	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	11.13	95	109279	44.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.60%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	3109	2.518	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	17221	8.417	ug/l	87
44) Trichloroethene	7.22	130	13332	7.169	ug/l	98
64) Tetrachloroethene	9.34	164	67699	42.333	ug/l	100
68) m/p-Xylenes	10.36	106	1450	0.462	ug/l	92
73) Isopropylbenzene	11.02	105	1897	0.275	ug/l	90
83) tert-Butylbenzene	11.77	119	3124	0.554	ug/l	88
91) 1,2-Dichlorobenzene	12.39	146	2246	0.724	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009698.D
Acq On : 18 May 2019 06:39
Operator : JC/SP
Sample : K2901-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-10B_R2

Quant Time: May 20 04:26:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

