

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009701.D
 Acq On : 18 May 2019 07:49
 Operator : JC/SP
 Sample : K2901-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-9B_R2

Quant Time: May 20 04:33:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	168230	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	269936	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	238626	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	99263	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	113787	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
35) Dibromofluoromethane	5.50	113	82630	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
50) Toluene-d8	8.71	98	338229	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
62) 4-Bromofluorobenzene	11.13	95	109987	44.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.36%	
Target Compounds						
64) Tetrachloroethene	9.34	164	789	0.482	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009701.D
Acq On : 18 May 2019 07:49
Operator : JC/SP
Sample : K2901-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-9B_R2

Quant Time: May 20 04:33:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

