

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009702.D
 Acq On : 18 May 2019 08:13
 Operator : JC/SP
 Sample : K2901-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15C_R2

Quant Time: May 20 04:36:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	168621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	270268	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	234791	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	95837	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115188	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
35) Dibromofluoromethane	5.50	113	83091	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	8.71	98	340660	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
62) 4-Bromofluorobenzene	11.13	95	109933	44.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.22%	
Target Compounds						
16) Acetone	2.45	43	2983	2.413	ug/l #	78
19) Methyl tert-butyl Ether	3.20	73	5599	0.911	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009702.D
Acq On : 18 May 2019 08:13
Operator : JC/SP
Sample : K2901-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
MW-15C_R2

Quant Time: May 20 04:36:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

