

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
 Data File : VX009707.D
 Acq On : 18 May 2019 10:10
 Operator : JC/SP
 Sample : K2901-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15A_R2

Quant Time: May 20 04:43:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	167152	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	269942	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	237636	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	102309	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	113313	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
35) Dibromofluoromethane	5.50	113	82247	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	8.71	98	338168	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
62) 4-Bromofluorobenzene	11.13	95	111487	44.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.56%	
Target Compounds						
16) Acetone	2.44	43	2408	1.965	ug/l	97
19) Methyl tert-butyl Ether	3.20	73	6213	1.020	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009707.D
Acq On : 18 May 2019 10:10
Operator : JC/SP
Sample : K2901-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 61 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15A_R2

Quant Time: May 20 04:43:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

