

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX051719\
 Data File : VX009710.D
 Acq On : 18 May 2019 11:20
 Operator : JC/SP
 Sample : K2901-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-04A_R2

Quant Time: May 25 07:20:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	170352	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	274624	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240437	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	100398	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116932	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
35) Dibromofluoromethane	5.50	113	85269	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
50) Toluene-d8	8.71	98	346846	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
62) 4-Bromofluorobenzene	11.13	95	112706	44.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.00%	
Target Compounds						
16) Acetone	2.45	43	3587	2.873	ug/l	98
21) trans-1,2-Dichloroethene	3.17	96	600	0.346	ug/l	97
83) tert-Butylbenzene	11.77	119	5015	0.839	ug/l	87
91) 1,2-Dichlorobenzene	12.39	146	886	0.269	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX051719\
Data File : VX009710.D
Acq On : 18 May 2019 11:20
Operator : JC/SP
Sample : K2901-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 64 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-04A_R2

Quant Time: May 25 07:20:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

