

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX051719\
 Data File : VX009689.D
 Acq On : 18 May 2019 02:45
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 18 06:47:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	71	0.00
2 T	Dichlorodifluoromethane	0.461	0.374	18.9	58	0.00
3 P	Chloromethane	0.629	0.592	5.9	69	0.00
4 C	Vinyl Chloride	0.613	0.584	4.7#	68	0.00
5 T	Bromomethane	0.368	0.371	-0.8	76	0.00
6 T	Chloroethane	0.394	0.379	3.8	70	0.00
7 T	Trichlorofluoromethane	0.746	0.733	1.7	71	0.00
8 T	Diethyl Ether	0.353	0.362	-2.5	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.442	5.6	68	0.00
10 T	Methyl Iodide	0.483	0.485	-0.4	67	0.00
11 T	Tert butyl alcohol	0.158	0.188	-19.0	86	0.01
12 CM	1,1-Dichloroethene	0.469	0.459	2.1#	69	0.00
13 T	Acrolein	0.143	0.140	2.1	71	0.00
14 T	Allyl chloride	1.143	1.114	2.5	71	0.00
15 T	Acrylonitrile	0.386	0.431	-11.7	83	0.00
16 T	Acetone	0.367	0.385	-4.9	78	0.00
17 T	Carbon Disulfide	1.317	1.160	11.9	62	0.00
18 T	Methyl Acetate	0.872	0.977	-12.0	84	0.00
19 T	Methyl tert-butyl Ether	1.822	1.929	-5.9	76	0.00
20 T	Methylene Chloride	0.593	0.579	2.4	73	0.00
21 T	trans-1,2-Dichloroethene	0.509	0.494	2.9	70	0.00
22 T	Diisopropyl ether	2.199	2.340	-6.4	77	0.00
23 T	Vinyl Acetate	1.883	2.053	-9.0	76	0.00
24 P	1,1-Dichloroethane	1.085	1.104	-1.8	75	0.00
25 T	2-Butanone	0.570	0.654	-14.7	82	0.00
26 T	2,2-Dichloropropane	0.807	0.527	34.7#	47#	0.00
27 T	cis-1,2-Dichloroethene	0.607	0.614	-1.2	74	0.00
28 T	Bromochloromethane	0.487	0.380	22.0#	55	0.00
29 T	Tetrahydrofuran	0.366	0.428	-16.9	84	0.00
30 C	Chloroform	0.981	1.022	-4.2#	76	0.00
31 T	Cyclohexane	1.009	0.957	5.2	67	0.00
32 T	1,1,1-Trichloroethane	0.817	0.832	-1.8	74	0.00
33 S	1,2-Dichloroethane-d4	0.683	0.660	3.4	68	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	71	0.00
35 S	Dibromofluoromethane	0.314	0.302	3.8	68	0.00
36 T	1,1-Dichloropropene	0.465	0.451	3.0	71	0.00
37 T	Ethyl Acetate	0.647	0.735	-13.6	82	0.00
38 T	Carbon Tetrachloride	0.422	0.422	0.0	73	0.00
39 T	Methylcyclohexane	0.571	0.513	10.2	65	0.00
40 TM	Benzene	1.438	1.441	-0.2	73	0.00
41 T	Methacrylonitrile	0.385	0.415	-7.8	80	0.00
42 TM	1,2-Dichloroethane	0.516	0.548	-6.2	79	0.00
43 T	Isopropyl Acetate	1.036	1.161	-12.1	81	0.00
44 TM	Trichloroethene	0.348	0.341	2.0	72	0.00
45 C	1,2-Dichloropropane	0.407	0.425	-4.4#	76	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX051719\
 Data File : VX009689.D
 Acq On : 18 May 2019 02:45
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 18 06:47:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.236	0.241	-2.1	76	0.00
47 T	Bromodichloromethane	0.472	0.494	-4.7	76	0.00
48 T	Methyl methacrylate	0.514	0.585	-13.8	82	0.00
49 T	1,4-Dioxane	0.010	0.011	-10.0	81	0.00
50 S	Toluene-d8	1.269	1.173	7.6	64	0.00
51 T	4-Methyl-2-Pentanone	0.677	0.807	-19.2	86	0.00
52 CM	Toluene	0.874	0.887	-1.5#	74	0.00
53 T	t-1,3-Dichloropropene	0.533	0.543	-1.9	72	0.00
54 T	cis-1,3-Dichloropropene	0.589	0.587	0.3	70	0.00
55 T	1,1,2-Trichloroethane	0.357	0.378	-5.9	77	0.00
56 T	Ethyl methacrylate	0.637	0.702	-10.2	78	0.00
57 T	1,3-Dichloropropane	0.628	0.660	-5.1	77	0.00
58 T	2-Chloroethyl Vinyl ether	0.338	0.293	13.3	60	0.00
59 T	2-Hexanone	0.532	0.624	-17.3	84	0.00
60 T	Dibromochloromethane	0.348	0.374	-7.5	76	0.00
61 T	1,2-Dibromoethane	0.360	0.381	-5.8	76	0.00
62 S	4-Bromofluorobenzene	0.461	0.447	3.0	68	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	74	0.00
64 T	Tetrachloroethene	0.343	0.350	-2.0	78	0.00
65 PM	Chlorobenzene	1.000	1.008	-0.8	76	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.374	-3.0	77	0.00
67 C	Ethyl Benzene	1.841	1.840	0.1#	75	0.00
68 T	m/p-Xylenes	0.673	0.672	0.1	74	0.00
69 T	o-Xylene	0.664	0.669	-0.8	76	0.00
70 T	Styrene	1.154	1.193	-3.4	76	0.00
71 P	Bromoform	0.295	0.316	-7.1	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
73 T	Isopropylbenzene	3.636	3.535	2.8	74	0.00
74 T	N-amyl acetate	2.120	2.265	-6.8	79	0.00
75 P	1,1,2,2-Tetrachloroethane	1.355	1.385	-2.2	79	0.00
76 T	1,2,3-Trichloropropane	1.160	1.212	-4.5	80	0.00
77 T	Bromobenzene	0.902	0.900	0.2	78	0.00
78 T	n-propylbenzene	4.152	4.084	1.6	74	0.00
79 T	2-Chlorotoluene	2.562	2.469	3.6	76	0.00
80 T	1,3,5-Trimethylbenzene	3.061	3.043	0.6	76	0.00
81 T	trans-1,4-Dichloro-2-butene	0.453	0.413	8.8	67	0.00
82 T	4-Chlorotoluene	2.894	2.868	0.9	76	0.00
83 T	tert-Butylbenzene	2.978	2.921	1.9	76	0.00
84 T	1,2,4-Trimethylbenzene	3.073	3.067	0.2	76	0.00
85 T	sec-Butylbenzene	3.531	3.445	2.4	74	0.00
86 T	p-Isopropyltoluene	3.174	3.125	1.5	74	0.00
87 T	1,3-Dichlorobenzene	1.609	1.579	1.9	75	0.00
88 T	1,4-Dichlorobenzene	1.622	1.582	2.5	75	0.00
89 T	n-Butylbenzene	2.869	2.779	3.1	72	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009689.D
Acq On : 18 May 2019 02:45
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 18 06:47:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.541	0.533	1.5	74	0.00
91 T	1,2-Dichlorobenzene	1.639	1.617	1.3	76	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.304	0.327	-7.6	80	0.00
93 T	1,2,4-Trichlorobenzene	1.121	1.121	0.0	74	0.00
94 T	Hexachlorobutadiene	0.572	0.553	3.3	73	0.00
95 T	Naphthalene	3.646	3.839	-5.3	78	0.00
96 T	1,2,3-Trichlorobenzene	1.136	1.147	-1.0	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6