


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051719\
Data File : VX009688.D
Acq On : 18 May 2019 01:58
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Title : SW846 8260
Last Update : Thu May 09 07:31:15 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	6108	PASS
75	95	30	60	49.9	14351	PASS
95	95	100	100	100.0	28760	PASS
96	95	5	9	6.2	1773	PASS
173	174	0.00	2	1.2	265	PASS
174	95	50	100	74.9	21550	PASS
175	174	5	9	8.7	1878	PASS
176	174	95	101	100.1	21567	PASS
177	176	5	9	6.6	1426	PASS