

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028761.D
 Acq On : 17 May 2022 16:32
 Operator : JC/MD
 Sample : N2809-04
 Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14A-TP-5R

Quant Time: May 17 16:59:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	310030	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	538486	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	492268	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	252503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	206372	48.561	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.120%
35) Dibromofluoromethane	5.391	113	168652	47.803	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.600%
50) Toluene-d8	8.647	98	738359	55.509	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.020%
62) 4-Bromofluorobenzene	11.085	95	255255	51.378	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.760%
Target Compounds						
					Qvalue	
25) 2-Butanone	4.605	43	24473	8.548	ug/l	93
39) Methylcyclohexane	7.373	83	125203	20.602	ug/l	93
67) Ethyl Benzene	10.195	91	37323	1.994	ug/l	96
69) o-Xylene	10.646	106	42134	5.872	ug/l	96
73) Isopropylbenzene	10.963	105	63629	3.323	ug/l	99
78) n-propylbenzene	11.305	91	102000	4.605	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	169171	10.603	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	112771	7.048	ug/l	100
85) sec-Butylbenzene	11.890	105	87697	4.602	ug/l #	76
86) p-Isopropyltoluene	12.012	119	72355	4.525	ug/l	96
89) n-Butylbenzene	12.329	91	115514	8.338	ug/l #	46

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028761.D
Acq On : 17 May 2022 16:32
Operator : JC/MD
Sample : N2809-04
Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14A-TP-5R

Quant Time: May 17 16:59:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

