

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028771.D
 Acq On : 17 May 2022 20:23
 Operator : JC/MD
 Sample : N2884-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31571

Quant Time: May 18 04:55:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	176022	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	6.769	114	173338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99011	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36920	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	30848	12.785	ug/l	0.01
Spiked Amount	50.000	Range	61 - 141	Recovery	=	25.560%#
35) Dibromofluoromethane	5.397	113	40837	35.958	ug/l	0.01
Spiked Amount	50.000	Range	69 - 133	Recovery	=	71.920%
50) Toluene-d8	8.653	98	181030	42.280	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	84.560%
62) 4-Bromofluorobenzene	11.085	95	40045	25.040	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	50.080%#
Target Compounds						
						Qvalue
16) Acetone	2.380	43	3237	3.126	ug/l	91
95) Naphthalene	13.780	128	4246	1.618	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028771.D
Acq On : 17 May 2022 20:23
Operator : JC/MD
Sample : N2884-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31571

Quant Time: May 18 04:55:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

