

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028773.D
 Acq On : 17 May 2022 21:10
 Operator : JC/MD
 Sample : N2879-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP1-0516

Quant Time: May 18 04:56:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	284313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	519538	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	555071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	233867	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	197852	50.767	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.540%
35) Dibromofluoromethane	5.391	113	170532	50.098	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.200%
50) Toluene-d8	8.653	98	621688	48.443	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.880%
62) 4-Bromofluorobenzene	11.079	95	306019	63.842	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	127.680%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	28715	17.168	ug/l	# 84
20) Methylene Chloride	2.788	84	19440	5.540	ug/l	94
43) Isopropyl Acetate	6.348	43	23022	2.756	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028773.D
Acq On : 17 May 2022 21:10
Operator : JC/MD
Sample : N2879-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP1-0516

Quant Time: May 18 04:56:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

