

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028774.D
 Acq On : 17 May 2022 21:33
 Operator : JC/MD
 Sample : N2879-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP2-0516

Quant Time: May 18 04:56:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	287207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	529949	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	609501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	289125	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	203220	51.619	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.240%
35) Dibromofluoromethane	5.391	113	171940	49.520	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.040%
50) Toluene-d8	8.653	98	635182	48.522	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.079	95	311015	63.609	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	127.220%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	37327	22.092	ug/l	89
20) Methylene Chloride	2.788	84	19203	5.417	ug/l	93
43) Isopropyl Acetate	6.349	43	24718	2.901	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028774.D
Acq On : 17 May 2022 21:33
Operator : JC/MD
Sample : N2879-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP2-0516

Quant Time: May 18 04:56:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

