

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028781.D
 Acq On : 18 May 2022 01:02
 Operator : JC/MD
 Sample : PB144875TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144875TB

Quant Time: May 18 04:59:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	349953	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	614749	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	593669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	282605	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	253076	52.757	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	105.520%	
35) Dibromofluoromethane	5.385	113	204044	50.660	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	101.320%	
50) Toluene-d8	8.653	98	742744	48.912	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	97.820%	
62) 4-Bromofluorobenzene	11.079	95	302453	53.325	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	106.660%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	30390	14.761	ug/l	93
20) Methylene Chloride	2.788	84	23777	5.505	ug/l	93
43) Isopropyl Acetate	6.348	43	31792	3.216	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028781.D
Acq On : 18 May 2022 01:02
Operator : JC/MD
Sample : PB144875TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144875TB

Quant Time: May 18 04:59:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

