

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028783.D
 Acq On : 18 May 2022 01:49
 Operator : JC/MD
 Sample : PB144875ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144875ZHE#02

Quant Time: May 18 05:00:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	349232	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	622779	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	597327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	285629	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	252764	52.800	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.600%
35) Dibromofluoromethane	5.385	113	206427	50.590	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.180%
50) Toluene-d8	8.653	98	741926	48.228	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.460%
62) 4-Bromofluorobenzene	11.079	95	304681	53.026	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.060%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	46684	22.722	ug/l	89
20) Methylene Chloride	2.788	84	18338	4.254	ug/l	88
43) Isopropyl Acetate	6.342	43	30065	3.002	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028783.D
Acq On : 18 May 2022 01:49
Operator : JC/MD
Sample : PB144875ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144875ZHE#02

Quant Time: May 18 05:00:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

