

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051722\  
 Data File : VX028785.D  
 Acq On : 18 May 2022 02:35  
 Operator : JC/MD  
 Sample : PB144875ZHE#04  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 41 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB144875ZHE#04

Quant Time: May 18 05:01:07 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051222W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 12 14:58:53 2022  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|------------|----------|
| -----                       |           |       |          |          |            |          |
| Internal Standards          |           |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.550     | 168   | 321102   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763     | 114   | 509207   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055    | 117   | 567431   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024    | 152   | 229062   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |           |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.958     | 65    | 235325   | 53.464   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 61 - 141 | Recovery | = 106.920% |          |
| 35) Dibromofluoromethane    | 5.385     | 113   | 166821   | 50.003   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 69 - 133 | Recovery | = 100.000% |          |
| 50) Toluene-d8              | 8.653     | 98    | 617032   | 49.055   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 65 - 126 | Recovery | = 98.120%  |          |
| 62) 4-Bromofluorobenzene    | 11.079    | 95    | 238218   | 50.705   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 58 - 135 | Recovery | = 101.420% |          |
| Target Compounds            |           |       |          |          |            | Qvalue   |
| 16) Acetone                 | 2.380     | 43    | 38668    | 20.470   | ug/l       | 91       |
| 20) Methylene Chloride      | 2.788     | 84    | 15472    | 3.904    | ug/l       | 91       |
| 43) Isopropyl Acetate       | 6.342     | 43    | 28082    | 3.430    | ug/l       | 97       |
| -----                       |           |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051722\  
Data File : VX028785.D  
Acq On : 18 May 2022 02:35  
Operator : JC/MD  
Sample : PB144875ZHE#04  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 41 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PB144875ZHE#04

Quant Time: May 18 05:01:07 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051222W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 12 14:58:53 2022  
Response via : Initial Calibration

