

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028786.D
 Acq On : 18 May 2022 02:59
 Operator : JC/MD
 Sample : PB144875ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144875ZHE#05

Quant Time: May 18 05:01:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	351361	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	625854	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	605816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	286092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	255251	52.997	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.000%	
35) Dibromofluoromethane	5.391	113	209178	51.013	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.020%	
50) Toluene-d8	8.653	98	747048	48.322	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.640%	
62) 4-Bromofluorobenzene	11.079	95	305491	52.905	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 105.820%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	44959	21.750	ug/l	95
20) Methylene Chloride	2.788	84	17932	4.135	ug/l	91
43) Isopropyl Acetate	6.342	43	30438	3.024	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028786.D
Acq On : 18 May 2022 02:59
Operator : JC/MD
Sample : PB144875ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144875ZHE#05

Quant Time: May 18 05:01:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

