

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028787.D
 Acq On : 18 May 2022 03:22
 Operator : JC/MD
 Sample : PB144875ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144875ZHE#06

Quant Time: May 18 05:01:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	334808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	599019	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	580755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	276089	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	242204	52.774	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.540%	
35) Dibromofluoromethane	5.391	113	199736	50.892	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.780%	
50) Toluene-d8	8.653	98	720979	48.725	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.460%	
62) 4-Bromofluorobenzene	11.079	95	292968	53.010	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 106.020%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	45258	22.977	ug/l	88
20) Methylene Chloride	2.788	84	18464	4.468	ug/l	95
43) Isopropyl Acetate	6.342	43	29214	3.033	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028787.D
Acq On : 18 May 2022 03:22
Operator : JC/MD
Sample : PB144875ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144875ZHE#06

Quant Time: May 18 05:01:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

