

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028788.D
 Acq On : 18 May 2022 03:45
 Operator : JC/MD
 Sample : N2879-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3-0516

Quant Time: May 18 05:02:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	336192	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	600314	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	575719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	271173	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	245576	53.289	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.580%	
35) Dibromofluoromethane	5.391	113	200021	50.855	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.700%	
50) Toluene-d8	8.653	98	715225	48.232	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.460%	
62) 4-Bromofluorobenzene	11.079	95	291018	52.543	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 105.080%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	35343	17.870	ug/l	88
20) Methylene Chloride	2.788	84	22660	5.461	ug/l	96
43) Isopropyl Acetate	6.348	43	25785	2.671	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028788.D
Acq On : 18 May 2022 03:45
Operator : JC/MD
Sample : N2879-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3-0516

Quant Time: May 18 05:02:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

