

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028790.D
 Acq On : 18 May 2022 04:32
 Operator : JC/MD
 Sample : N2881-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SLP-1

Quant Time: May 18 07:16:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	341179	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	606911	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	577665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	267657	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	249948	53.445	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 106.880%		
35) Dibromofluoromethane	5.385	113	200493	50.421	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 100.840%		
50) Toluene-d8	8.653	98	726230	48.442	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 96.880%		
62) 4-Bromofluorobenzene	11.079	95	285999	51.076	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 102.160%		
Target Compounds						
						Qvalue
16) Acetone	2.380	43	48082	23.955	ug/l	88
20) Methylene Chloride	2.788	84	21205	5.036	ug/l	# 89
43) Isopropyl Acetate	6.342	43	27588	2.827	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028790.D
Acq On : 18 May 2022 04:32
Operator : JC/MD
Sample : N2881-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SLP-1

Quant Time: May 18 07:16:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

