

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028791.D
 Acq On : 18 May 2022 04:55
 Operator : JC/MD
 Sample : N2882-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SLP-2

Quant Time: May 18 07:16:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	330144	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	591104	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	544813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	247614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	240678	53.183	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.360%	
35) Dibromofluoromethane	5.391	113	196822	50.821	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.640%	
50) Toluene-d8	8.653	98	693678	47.508	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.020%	
62) 4-Bromofluorobenzene	11.079	95	263887	48.387	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.780%	
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	65314	33.628	ug/l	92
20) Methylene Chloride	2.788	84	25285	6.205	ug/l	96
25) 2-Butanone	4.562	43	16235	5.325	ug/l	95
43) Isopropyl Acetate	6.348	43	21475	2.259	ug/l	95
52) Toluene	8.720	92	20730	1.928	ug/l	94
68) m/p-Xylenes	10.305	106	8510	1.055	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028791.D
Acq On : 18 May 2022 04:55
Operator : JC/MD
Sample : N2882-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SLP-2

Quant Time: May 18 07:16:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

