

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
 Data File : VX028792.D
 Acq On : 18 May 2022 05:19
 Operator : JC/MD
 Sample : N2858-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-38

Quant Time: May 18 07:17:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	290914	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	562771	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	456433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	235189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	206183	51.704	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.400%
35) Dibromofluoromethane	5.391	113	175819	47.684	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.360%
50) Toluene-d8	8.652	98	623545	44.855	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	89.700%
62) 4-Bromofluorobenzene	11.079	95	218890	42.157	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.320%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	44471	25.984	ug/l	92
20) Methylene Chloride	2.788	84	99558	27.728	ug/l	91
43) Isopropyl Acetate	6.348	43	23472	2.594	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051722\
Data File : VX028792.D
Acq On : 18 May 2022 05:19
Operator : JC/MD
Sample : N2858-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-38

Quant Time: May 18 07:17:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

