

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
 Data File : VX035706.D
 Acq On : 17 May 2023 10:29
 Operator : JC/MD
 Sample : VX0517WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0517WBL01

Quant Time: May 18 04:38:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	182510	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	338180	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	307932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	136878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	154341	48.475	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.940%
35) Dibromofluoromethane	5.385	113	111028	48.682	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.360%
50) Toluene-d8	8.647	98	407700	48.389	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.780%
62) 4-Bromofluorobenzene	11.079	95	165571	48.806	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
Data File : VX035706.D
Acq On : 17 May 2023 10:29
Operator : JC/MD
Sample : VX0517WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0517WBL01

Quant Time: May 18 04:38:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

