

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
 Data File : VX035711.D
 Acq On : 17 May 2023 12:28
 Operator : JC/MD
 Sample : 02816-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-051623

Quant Time: May 18 04:40:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	185522	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	345382	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	322082	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146311	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162274	50.139	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.280%
35) Dibromofluoromethane	5.385	113	112928	48.483	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.960%
50) Toluene-d8	8.647	98	416830	48.441	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.880%
62) 4-Bromofluorobenzene	11.079	95	174977	50.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051723\
Data File : VX035711.D
Acq On : 17 May 2023 12:28
Operator : JC/MD
Sample : 02816-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-051623

Quant Time: May 18 04:40:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

